# Store at +4C

# PathScan<sup>®</sup> RP Phospho-Aurora A (Thr288) Sandwich ELISA Kit



Orders: 877-616-CELL (2355)

orders@cellsignal.com

Support: 877-678-TECH (8324)

Web: info@cellsignal.com

cellsignal.com

3 Trask Lane | Danvers | Massachusetts | 01923 | USA

1 Kit (96 assays)

Species Cross Reactivity: UniProt ID: Entrez-Gene Id: #014965 #6790

For Research Use Only. Not for Use in Diagnostic Procedures.

Product Includes	Product #	Quantity	Color	Storage Temp
Phospho-Aurora A (Thr288) Rabbit mAb Coated Microwells	52359	96 tests		+4C
Aurora A Rabbit Detection mAb	62749	1 ea	Red (Lyophilized)	+4C
HRP Diluent	13515	5.5 ml	Red	+4C
TMB Substrate	7004	11 ml		+4C
STOP Solution	7002	11 ml		+4C
Sealing Tape	54503	2 ea		+4C
ELISA Wash Buffer (20X)	9801	25 ml		+4C
Cell Lysis Buffer (10X)	9803	15 ml		-20C

Kit contents scale proportionally with size, except sealing tape.

Example: The V1 kit contains 5X the listed quantities above, but will exclude the sealing tape.

The microwell plate is supplied as 12 8-well modules - Each module is designed to break apart for 8 tests.

# Description

The rapid protocol (RP) PathScan<sup>®</sup> RP Phospho-Aurora A (Thr288) Sandwich ELISA Kit is a solid phase sandwich enzyme-linked immunosorbent assay (ELISA) that detects endogenous levels of Aurora A protein phosphorylated at Thr288 in a reduced assay time of 1.5 hours. Incubation of cell lysates and detection antibody on the coated microwell plate forms a sandwich with phospho-Aurora A (Thr288) in a single step. The plate is then extensively washed and TMB reagent is added for signal development. The magnitude of absorbance for the developed color is proportional to the quantity of phospho-Aurora A (Thr288). Learn more about your ELISA kit options here.

# Specificity/Sensitivity

The PathScan® RP Phospho Aurora A (Thr288) Sandwich ELISA Kit detects endogenous levels of Aurora A protein phosphorylated at Thr288. The kit sensitivity is shown in Figure 1. This kit detects proteins from the indicated species, as determined through in-house testing, but may also detect homologous proteins from other species.

# **Background**

Aurora kinases belong to a highly conserved family of mitotic serine/threonine kinases with three members identified among mammals: Aurora A, B, and C (1,2). Studies on the temporal expression pattern and subcellular localization of Aurora kinases in mitotic cells suggest an association with mitotic structure. Aurora kinase functional influences span from G2 phase to cytokinesis and may be involved in key cell cycle events such as centrosome duplication, chromosome bi-orientation and segregation, cleavage furrow positioning, and ingression (3). Aurora A is detected at the centrosomes, along mitotic spindle microtubules, and in the cytoplasm of mitotically proliferating cells. Aurora A protein levels are low during G1 and S phases and peak during the G2/M phase of the cell cycle. Phosphorylation of Aurora A at Thr288 in its catalytic domain increases kinase activity. Aurora A is involved in centrosome separation, maturation, and spindle assembly and stability. Expression of Aurora B protein also peaks during the G2/M phase of the cell cycle; Aurora B kinase activity peaks at the transition from metaphase to the end of mitosis. Aurora B associates with chromosomes during prophase prior to relocalizing to the spindle at anaphase. Aurora B regulates chromosome segregation through the control of microtubule-kinetochore attachment and cytokinesis. Expression of both Aurora A and Aurora B during the G2/M phase transition is tightly coordinated with histone H3 phosphorylation (4,5); research investigators have observed overexpression of these kinases in a variety of human cancers (2,4). Aurora C localizes to the centrosome from anaphase to cytokinesis and both mRNA and protein levels peak during G2/M phase. Although typical Aurora C expression is limited to the testis, research studies report overexpression of Aurora C is detected in various cancer cell lines

# **Background References**

<sup>\*</sup>Antibodies in this kit are custom formulations specific to kit.

- 2. Katayama, H. et al. (2003) Cancer Metastasis Rev 22, 451-64.
- 3. Andrews, P.D. et al. (2003) Curr Opin Cell Biol 15, 672-83.
- 4. Pascreau, G. et al. (2003) Prog Cell Cycle Res 5, 369-74.
- 5. Crosio, C. et al. (2002) Mol Cell Biol 22, 874-85.
- 6. Kimura, M. et al. (1999) / Biol Chem 274, 7334-40.

#### Trademarks and Patents

Cell Signaling Technology is a trademark of Cell Signaling Technology, Inc.

PathScan is a registered trademark of Cell Signaling Technology, Inc.

U.S. Patent No. 7,429,487, foreign equivalents, and child patents deriving therefrom.

All other trademarks are the property of their respective owners. Visit cellsignal.com/trademarks for more information.

#### **Limited Uses**

Except as otherwise expressly agreed in a writing signed by a legally authorized representative of CST, the following terms apply to Products provided by CST, its affiliates or its distributors. Any Customer's terms and conditions that are in addition to, or different from, those contained herein, unless separately accepted in writing by a legally authorized representative of CST, are rejected and are of no force or effect.

Products are labeled with For Research Use Only or a similar labeling statement and have not been approved, cleared, or licensed by the FDA or other regulatory foreign or domestic entity, for any purpose. Customer shall not use any Product for any diagnostic or therapeutic purpose, or otherwise in any manner that conflicts with its labeling statement. Products sold or licensed by CST are provided for Customer as the end-user and solely for research and development uses. Any use of Product for diagnostic, prophylactic or therapeutic purposes, or any purchase of Product for resale (alone or as a component) or other commercial purpose, requires a separate license from CST. Customer shall (a) not sell, license, loan, donate or otherwise transfer or make available any Product to any third party, whether alone or in combination with other materials, or use the Products to manufacture any commercial products, (b) not copy, modify, reverse engineer, decompile, disassemble or otherwise attempt to discover the underlying structure or technology of the Products, or use the Products for the purpose of developing any products or services that would compete with CST products or services, (c) not alter or remove from the Products any trademarks, trade names, logos, patent or copyright notices or markings, (d) use the Products solely in accordance with CST Product Terms of Sale and any applicable documentation, and (e) comply with any license, terms of service or similar agreement with respect to any third party products or services used by Customer in connection with the Products.

# #41333

# PathScan<sup>®</sup> RP Phospho-Aurora A (Thr288) Sandwich ELISA Kit



# PathScan® Sandwich ELISA Protocol (Rapid Protocol)

**NOTE:** This protocol is for PathScan<sup>®</sup> kits that use an HRP directly conjugated to the detection antibody **(Rapid Protocol)**, rather than a 2-step method where the detection antibody and a secondary-HRP are added sequentially.

# A. Solutions and Reagents

**NOTE**: Prepare solutions with deionized/purified water or equivalent.

- 1. **Microwell strips:** Bring all to room temperature before opening bag/use. Unused microwell strips should be returned to the original re-sealable bag containing the desiccant pack and stored at 4°C.
- 2. **Detection Antibody:** Reconstitute lyophilized Detection Antibody (red colored cake) with 1 mL of HRP Diluent (red solution) to yield a concentrated stock solution. Incubate at room temperature for 5 min with occasional gentle mixing to fully reconstitute. To make the final working solution, add the full 1 mL of reconstituted Detection Antibody to 4.5 mL of HRP Diluent in a clean tube and gently mix. For best results, use immediately following antibody reconstitution. Unused reconstituted Detection Antibody may be stored for up to 4 weeks at 4°C, although there may be some loss of signal compared to freshly reconstituted antibody.
- 3. HRP Diluent: Red colored diluent for reconstitution and dilution of the Detection Antibody that is linked to HRP.
- 4. **1X ELISA Wash Buffer:** Prepare by diluting ELISA Wash Buffer (20X) (included in each kit) to 1X with deionized water.
- 5. **1X Cell Lysis Buffer:** Prepare by diluting 10X Cell Lysis Buffer #9803 to 1X with deionized water. This buffer can be stored at 4°C for short-term use (1–2 weeks). Recommended: When using to prepare cell lysates, add Protease/Phosphatase Inhibitor Cocktail (#5872, not supplied) and 1 mM phenylmethyl- sulfonyl fluoride (PMSF, #8553, not supplied) immediately before use.
- 6. **TMB Substrate (#7004):** Bring to room temperature before use.
- 7. **STOP Solution (#7002):** Bring to room temperature before use.

# **B. Preparing Cell Lysates**

### For adherent cells

- 1. Aspirate media when the culture reaches 80–90% confluence. Treat cells by adding fresh media containing regulator for desired time.
- 2. Remove media and rinse cells once with ice-cold 1X PBS.
- 3. Remove PBS and add 0.5 mL ice-cold 1X Cell Lysis Buffer including 1 mM PMSF and Protease/Phosphatase Inhibitor Cocktail to each plate (10 cm diameter) and incubate the plate on ice for 5 min.
- 4. Scrape cells off the plate and transfer to an appropriate tube. Keep on ice.
- 5. Sonicate lysates on ice.
- 6. Microcentrifuge for 10 min (14,000 rpm) at 4°C and transfer the supernatant to a new tube. The supernatant is the cell lysate. Store at -80°C in single-use aliquots.

### For suspension cells

- 1. Remove media by low speed centrifugation ( $\sim$ 1200 rpm) when the culture reaches 0.5–1.0 x 10 $^6$  viable cells/mL. Treat cells by adding fresh media containing regulator for desired time.
- 2. Collect cells by low speed centrifugation (~1200 rpm) and wash once with 5-10 mL ice-cold 1X PBS.
- 3. Cells harvested from 50 mL of growth media can be lysed in 2.0 mL of 1X Cell Lysis Buffer including 1 mM PMSF and Protease/Phosphatase Inhibitor Cocktail.
- 4. Sonicate lysates on ice.
- 5. Microcentrifuge for 10 min (14,000 rpm) at 4°C and transfer the supernatant to a new tube. The supernatant is the cell lysate. Store at -80°C in single-use aliquots.

## C. Test Procedure

NOTE: Equilibrate all materials and prepared reagents to room temperature prior to running the assay.

- 1. Prepare all reagents as indicated above (Section A).
- 2. Samples should be undiluted or diluted with 1X Cell Lysis Buffer to a 2X protein concentration in order to achieve a final 1X protein concentration upon addition of the Detection Antibody. Individual datasheets for each kit provide a sensitivity curve that serves as a reference for selection of an appropriate starting lysate concentration. The sensitivity curve shows typical results across a range of lysate concentration points.
- 3. Add 50  $\mu$ L of each sample to the appropriate wells.
- 4. Add 50  $\mu\text{L}$  of the Detection Antibody to each well.
- 5. Seal the plate and incubate for 1 hour at room temperature on a plate shaker set to 400 rpm (moderate agitation).

- 6. Gently remove the tape and wash wells:
  - 1. Discard plate contents into a receptacle.
  - 2. Wash 4 times with 1X Wash Buffer, 200  $\mu L$  each time for each well.
  - 3. For each wash, strike plates on fresh towels hard enough to remove the residual solution in each well, but do not allow wells to completely dry at any time.
  - 4. Clean the underside of all wells with a lint-free tissue.
- 7. Add 100 µL of TMB Substrate to each well. Seal with tape and incubate the plate in the dark for 15 min at room temperature on a plate shaker (400 rpm, moderate agitation) or alternatively for 10 min at 37°C without shaking.
- 8. Add 100  $\mu L$  of STOP Solution to each well. Shake gently for a few seconds.

NOTE: Initial color of positive reaction is blue, which changes to yellow upon addition of STOP Solution.

- 9. Read results:
  - 1. **Visual Determination:** Read within 30 min after adding STOP Solution.
  - 2. **Spectrophotometric Determination:** Wipe underside of wells with a lint-free tissue. Read absorbance at 450 nm within 30 min after adding STOP Solution.

created July 2020