

PathScan® Acetylated Histone H3 Sandwich ELISA Antibody Pair

Orders: 877-616-CELL (2355)
orders@cellsignal.com

Support: 877-678-TECH (8324)

Web: info@cellsignal.com
cellsignal.com

Species Cross Reactivity: H M Mk
UniProt ID: #P68431
Entrez-Gene Id: #8350

3 Trask Lane | Danvers | Massachusetts | 01923 | USA

For Research Use Only. Not for Use in Diagnostic Procedures.

Description

CST's PathScan® Acetylated Histone H3 Sandwich ELISA Antibody Pair is being offered as an alternative to our PathScan® Acetylated Histone H3 Sandwich ELISA Kit #7232. Capture and Detection antibodies (100X stocks) and HRP-conjugated secondary antibody (1000X stock) are supplied. Sufficient reagents are provided for performing 4 x 96 well ELISAs. The Histone Capture Antibody is coated in PBS overnight in a 96 well microplate. After blocking, cell lysates are added, followed by an Acetylated-Lysine Detection Antibody and HRP-conjugated Anti-Mouse IgG antibody. HRP substrate, TMB, is added for color development. The magnitude of the absorbance at 450 nm is proportional to the quantity of Acetylated Histone H3 protein.

*Antibodies in this kit are custom formulations specific to the kit

Reagents Not Supplied

Phosphate Buffered Saline (PBS-20X) #9808 Phosphate Buffered Saline with Tween -20 (PBST-20X) #9809 Cell Lysis Buffer (10X) #9803 TMB Substrate #7004 STOP Solution #7002 Blocking Buffer-PBS+0.05% Tween-20, 1% BSA 96 Well Microplates** Microplate Reader ** Antibody Pairs have been validated on Corning® 96 Well Clear Polystyrene High Bind Stripwell™ Microplates (#2592) and Corning® 96 Well EIA/RIA Easy Wash™ Clear Flat Bottom Polystyrene High Bind Microplates (#3369).

Background

Modulation of chromatin structure plays an important role in the regulation of transcription in eukaryotes. The nucleosome, made up of DNA wound around eight core histone proteins (two each of H2A, H2B, H3, and H4), is the primary building block of chromatin (1). The amino-terminal tails of core histones undergo various posttranslational modifications, including acetylation, phosphorylation, methylation, and ubiquitination (2-5). These modifications occur in response to various stimuli and have a direct effect on the accessibility of chromatin to transcription factors and, therefore, gene expression (6). In most species, histone H2B is primarily acetylated at Lys5, 12, 15, and 20 (4,7). Histone H3 is primarily acetylated at Lys9, 14, 18, 23, 27, and 56. Acetylation of H3 at Lys9 appears to have a dominant role in histone deposition and chromatin assembly in some organisms (2,3). Phosphorylation at Ser10, Ser28, and Thr11 of histone H3 is tightly correlated with chromosome condensation during both mitosis and meiosis (8-10). Phosphorylation at Thr3 of histone H3 is highly conserved among many species and is catalyzed by the kinase haspin. Immunostaining with phospho-specific antibodies in mammalian cells reveals mitotic phosphorylation at Thr3 of H3 in prophase and its dephosphorylation during anaphase (11).

Background References

1. Workman, J.L. and Kingston, R.E. (1998) *Annu Rev Biochem* 67, 545-79.
2. Hansen, J.C. et al. (1998) *Biochemistry* 37, 17637-41.
3. Strahl, B.D. and Allis, C.D. (2000) *Nature* 403, 41-5.
4. Cheung, P. et al. (2000) *Cell* 103, 263-71.
5. Bernstein, B.E. and Schreiber, S.L. (2002) *Chem Biol* 9, 1167-73.
6. Jaskelioff, M. and Peterson, C.L. (2003) *Nat Cell Biol* 5, 395-9.
7. Thorne, A.W. et al. (1990) *Eur J Biochem* 193, 701-13.
8. Hendzel, M.J. et al. (1997) *Chromosoma* 106, 348-60.
9. Goto, H. et al. (1999) *J Biol Chem* 274, 25543-9.
10. Preuss, U. et al. (2003) *Nucleic Acids Res* 31, 878-85.
11. Dai, J. et al. (2005) *Genes Dev* 19, 472-88.

Trademarks and Patents

Cell Signaling Technology is a trademark of Cell Signaling Technology, Inc.

PathScan is a registered trademark of Cell Signaling Technology, Inc.

All other trademarks are the property of their respective owners. Visit cellsignal.com/trademarks for more information.

Limited Uses

Except as otherwise expressly agreed in a writing signed by a legally authorized representative of CST, the following terms apply to Products provided by CST, its affiliates or its distributors. Any Customer's terms and conditions that are in addition to, or different from, those contained herein, unless separately accepted in writing by a legally authorized representative of CST, are rejected and are of no force or effect.

Products are labeled with For Research Use Only or a similar labeling statement and have not been approved, cleared, or licensed by the FDA or other regulatory foreign or domestic entity, for any purpose. Customer shall not use any Product for any diagnostic or therapeutic purpose, or otherwise in any manner that conflicts with its labeling statement. Products sold or licensed by CST are provided for Customer as the end-user and solely for research and development uses. Any use of Product for diagnostic, prophylactic or therapeutic purposes, or any purchase of Product for resale (alone or as a component) or other commercial purpose, requires a separate license from CST. Customer shall (a) not sell, license, loan, donate or otherwise transfer or make available any Product to any third party, whether alone or in combination with other materials, or use the Products to manufacture any commercial products, (b) not copy, modify, reverse engineer, decompile, disassemble or otherwise attempt to discover the underlying structure or technology of the Products, or use the Products for the purpose of developing any products or services that would compete with CST products or services, (c) not alter or remove from the Products any trademarks, trade names, logos, patent or copyright notices or markings, (d) use the Products solely in accordance with CST Product Terms of Sale and any applicable documentation, and (e) comply with any license, terms of service or similar agreement with respect to any third party products or services used by Customer in connection with the Products.

#7209

PathScan® Acetylated Histone H3 Sandwich ELISA Antibody Pair



ELISA Antibody Pair

A. Solutions and Reagents

NOTE: Prepare solutions with reverse osmosis deionized (RODI) or equivalent grade water.

1. **20X Phosphate Buffered Saline (PBS):** (#9808) To prepare 1 L 1X PBS: add 50 ml 20X PBS to 950 ml dH₂O, mix.
2. **Wash Buffer:** 1X PBS/0.05% Tween® 20, (20X PBST #9809).
3. **Blocking Buffer:** 1X PBS/0.05% Tween® 20, 1% BSA.
4. **1X Cell Lysis Buffer:** 10X Cell Lysis Buffer (#9803): To prepare 10 ml of 1X Cell Lysis Buffer, add 1 ml of 10X Cell Lysis Buffer to 9 ml of dH₂O, mix. Buffer can be stored at 4°C for short-term use (1–2 weeks).

Recommended: Add 1 mM phenylmethylsulfonyl fluoride (PMSF) (#8553) immediately before use.

5. **Bovine Serum Albumin (BSA):** (#9998).
6. **TMB Substrate:** (#7004).
7. **STOP Solution:** (#7002)

NOTE: Reagents should be made fresh daily.

B. Preparing Cell Lysates

For adherent cells

1. Aspirate media when the culture reaches 80–90% confluence. Treat cells by adding fresh media containing regulator for desired time.
2. Remove media and rinse cells once with ice-cold 1X PBS.
3. Remove PBS and add 0.5 ml ice-cold 1X Cell Lysis Buffer plus 1 mM PMSF to each plate (10 cm diameter) and incubate the plate on ice for 5 min.
4. Scrape cells off the plate and transfer to an appropriate tube. Keep on ice.
5. Sonicate lysates on ice.
6. Microcentrifuge for 10 min (x14,000 rpm) at 4°C and transfer the supernatant to a new tube. The supernatant is the cell lysate. Store at -80°C in single-use aliquots.

For suspension cells

1. Remove media by low speed centrifugation (~1,200 rpm) when the culture reaches 0.5–1.0 x 10⁶ viable cells/ml. Treat cells by adding fresh media containing regulator for desired time.
2. Collect cells by low speed centrifugation (~1,200 rpm) and wash once with 5–10 ml ice-cold 1X PBS.
3. Cells harvested from 50 ml of growth media can be lysed in 2.0 ml of 1X cell lysis buffer plus 1 mM PMSF.
4. Sonicate lysates on ice.
5. Microcentrifuge for 10 min (x14,000 rpm) at 4°C and transfer the supernatant to a new tube. The supernatant is the cell lysate. Store at -80°C in single-use aliquots.

C. Coating Procedure

1. Rinse microplate with 200 µl of dH₂O, discard liquid. Blot on paper towel to make sure wells are dry.
2. Dilute capture antibody 1:100 in 1X PBS. For a single 96 well plate, add 100 µl of capture antibody stock to 9.9 ml 1X PBS. Mix well and add 100 µl/well. Cover plate and incubate overnight at 4°C (17–20 hr).
3. **After overnight coating, gently uncover plate and wash wells:**
 1. Discard plate contents into a receptacle.
 2. Wash four times with wash buffer, 200 µl each time per well. For each wash, strike plates on fresh paper towels hard enough to remove the residual solution in each well, but do not allow wells to completely dry at any time.
 3. Clean the underside of all wells with a lint-free tissue.
4. Block plates. Add 150 µl of blocking buffer/well, cover plate, and incubate at 37°C for 2 hr.
5. After blocking, wash plate (Section C, Step 3). Plate is ready to use.

D. Test Procedure

1. Lysates can be used undiluted or diluted in blocking buffer. 100 µl of lysate is added per well. Cover plate and incubate at 37°C for 2 hr.
2. Wash plate (Section C, Step 3).
3. Dilute detection antibody 1:100 in blocking buffer. For a single 96 well plate, add 100 µl of detection antibody Stock to 9.9 ml of blocking buffer. Mix well and add 100 µl/well. Cover plate and incubate at 37°C for 1 hr.
4. Wash plate (Section C, Step 3).
5. Secondary antibody, either streptavidin anti-mouse or anti-rabbit-HRP, is diluted 1:1000 in blocking buffer. For a single 96 well plate, add 10 µl of secondary antibody stock to 9.99 ml of blocking buffer. Mix well and add 100 µl/well. Cover and incubate at 37°C for 30 min.
6. Wash plate (Section C, Step 3).
7. Add 100 µl of TMB substrate per well. Cover and incubate at 37°C for 10 min.
8. Add 100 µl of STOP solution per well. Shake gently for a few seconds.
9. Read plate on a microplate reader at absorbance 450 nm.
 1. **Visual Determination:** Read within 30 min after adding STOP solution.
 2. **Spectrophotometric Determination:** Wipe underside of wells with a lint-free tissue. Read absorbance at 450 nm within 30 min after adding STOP solution.

posted January 2008

revised September 2013