PathScan[®] Phospho-VEGFR-2 (Tyr1175) Sandwich ELISA Antibody Pair



Orders: 877-616-CELL (2355) orders@cellsignal.com

877-678-TECH (8324)

Support:

Web: info@cellsignal.com

cellsignal.com

3 Trask Lane | Danvers | Massachusetts | 01923 | USA

Species Cross Reactivity:

UniProt ID: Entrez-Gene Id:

For Research Use Only. Not for Use in Diagnostic Procedures.

| Product Includes | Product # | Volume | Cap Color | Storage Temp |
|---|-----------|--------|-----------|--------------|
| VEGFR-2 Capture Mouse mAb (100X) | 86047 | 400 µl | Pink | +4C |
| VEGFR-2 (Tyr1175) Detection Rabbit mAb (100X) | 98504 | 400 µl | Blue | +4C |
| Anti-rabbit IgG, HRP-linked Antibody (1000X) | 25944 | 40 µl | Red | -20C |

Please visit cellsignal.com for a complete listing of recommended companion products.

Description

CST's PathScan® Phospho-VEGFR-2 (Tyr1175) Sandwich ELISA Antibody Pair is offered as an economical alternative to our PathScan® Phospho-VEGFR-2 (Tyr1175) Sandwich ELISA Kit #7335. Capture and detection antibodies (100X stocks) and HRP-conjugated secondary antibody (1000X stock) are supplied. Sufficient reagents are supplied for 4 x 96 well ELISAs. The VEGFR-2 capture antibody is coated in PBS overnight in a 96 well microplate. After blocking, cell lysates are added followed by a Phospho-VEGFR-2 (Tyr1175) detection antibody and anti-rabbit IgG, HRP conjugated antibody. HRP substrate (TMB) is added for color development. The magnitude of the absorbance for this developed color is proportional to the quantity of Phospho-VEGFR-2 (Tyr1175) protein.

Antibodies in kit are custom formulations specific to kit.

Reagents Not Supplied

Phosphate Buffered Saline (PBS-20X) #9808

Phosphate Buffered Saline with Tween-20 (PBST-20X) #9809

Cell Lysis Buffer (10X) #9803 TMB Substrate #7004 STOP Solution #7002

Blocking Buffer: 1X PBS/0.5% Tween-20, 1% BSA

96 Well Microplates** Microplate Reader

** Antibody Pairs have been validated on Corning© 96 Well Clear Polystyrene High Bind Stripwell™

Microplates (#2592).

Notes: Antibody pairs have been optimized using recommended buffers, reagents, plates and the included protocol. Solutions should be made fresh daily.

Background

Vascular endothelial growth factor receptor 2 (VEGFR2, KDR, Flk-1) is a major receptor for VEGF-induced signaling in endothelial cells. Upon ligand binding, VEGFR2 undergoes autophosphorylation and becomes activated (1). Major autophosphorylation sites of VEGFR2 are located in the kinase insert domain (Tyr951/996) and in the tyrosine kinase catalytic domain (Tyr1054/1059) (2). Activation of the receptor leads to rapid recruitment of adaptor proteins, including Shc, GRB2, PI3 kinase, NCK, and the protein tyrosine phosphatases SHP-1 and SHP-2 (3). Phosphorylation at Tyr1212 provides a docking site for GRB2 binding and phospho-Tyr1175 binds the p85 subunit of PI3 kinase and PLCγ, as well as Shb (1,4,5). Signaling from VEGFR2 is necessary for the execution of VEGF-stimulated proliferation, chemotaxis and sprouting, as well as survival of cultured endothelial cells *in vitro* and angiogenesis *in* vivo (6-8).

Background References

- 1. Meyer, M. et al. (1999) EMBO J 18, 363-74.
- 2. Dougher-Vermazen, M. et al. (1994) Biochem Biophys Res Commun 205, 728-38.
- 3. Kroll, J. and Waltenberger, J. (1997) J Biol Chem 272, 32521-7.
- 4. Takahashi, T. et al. (2001) EMBO J 20, 2768-78.
- 5. Holmqvist, K. et al. (2004) J Biol Chem 279, 22267-75.
- 6. Karkkainen, M.J. and Petrova, T.V. (2000) Oncogene 19, 5598-605.
- 7. Rahimi, N. et al. (2000) *J Biol Chem* 275, 16986-92.
- 8. Claesson-Welsh, L. (2003) Biochem Soc Trans 31, 20-4.

Trademarks and Patents

Cell Signaling Technology is a trademark of Cell Signaling Technology, Inc.

PathScan is a registered trademark of Cell Signaling Technology, Inc.

All other trademarks are the property of their respective owners. Visit cellsignal.com/trademarks for more information.

Limited Uses

Except as otherwise expressly agreed in a writing signed by a legally authorized representative of CST, the following terms apply to Products provided by CST, its affiliates or its distributors. Any Customer's terms and conditions that are in addition to, or different from, those contained herein, unless separately accepted in writing by a legally authorized representative of CST, are rejected and are of no force or effect.

Products are labeled with For Research Use Only or a similar labeling statement and have not been approved, cleared, or licensed by the FDA or other regulatory foreign or domestic entity, for any purpose. Customer shall not use any Product for any diagnostic or therapeutic purpose, or otherwise in any manner that conflicts with its labeling statement. Products sold or licensed by CST are provided for Customer as the end-user and solely for research and development uses. Any use of Product for diagnostic, prophylactic or therapeutic purposes, or any purchase of Product for resale (alone or as a component) or other commercial purpose, requires a separate license from CST. Customer shall (a) not sell, license, loan, donate or otherwise transfer or make available any Product to any third party, whether alone or in combination with other materials, or use the Products to manufacture any commercial products, (b) not copy, modify, reverse engineer, decompile, disassemble or otherwise attempt to discover the underlying structure or technology of the Products, or use the Products for the purpose of developing any products or services that would compete with CST products or services, (c) not alter or remove from the Products any trademarks, trade names, logos, patent or copyright notices or markings, (d) use the Products solely in accordance with CST Product Terms of Sale and any applicable documentation, and (e) comply with any license, terms of service or similar agreement with respect to any third party products or services used by Customer in connection with the Products.

Orders: 877-616-CELL (2355) • orders@cellsignal.com • Support: 877-678-TECH (8324) • info@cellsignal.com • Web: cellsignal.com For Research Use Only. Not for Use in Diagnostic Procedures.

#7824

PathScan[®] Phospho-VEGFR-2 (Tyr1175) Sandwich ELISA Antibody Pair



ELISA Antibody Pair

A. Solutions and Reagents

NOTE: Prepare solutions with reverse osmosis deionized (RODI) or equivalent grade water.

- 20X Phosphate Buffered Saline (PBS): (#9808) To prepare 1 L 1X PBS: add 50 ml 20X PBS to 950 ml dH₂O, mix.
- 2. **Wash Buffer**: 1X PBS/0.05% Tween[®] 20, (20X PBST #9809).
- 3. Blocking Buffer: 1X PBS/0.05% Tween® 20, 1% BSA.
- 4. **1X Cell Lysis Buffer**: 10X Cell Lysis Buffer (#9803): To prepare 10 ml of 1X Cell Lysis Buffer, add 1 ml of 10X Cell Lysis Buffer to 9 ml of dH₂O, mix. Buffer can be stored at 4°C for short-term use (1–2 weeks).

Recommended: Add 1 mM phenylmethylsulfonyl fluoride (PMSF) (#8553) immediately before use.

- 5. Bovine Serum Albumin (BSA): (#9998).
- 6. **TMB Substrate**: (#7004).
- 7. **STOP Solution**: (#7002)

NOTE: Reagents should be made fresh daily.

B. Preparing Cell Lysates

For adherent cells

- Aspirate media when the culture reaches 80-90% confluence. Treat cells by adding fresh media containing regulator for desired time.
- 2. Remove media and rinse cells once with ice-cold 1X PBS.
- 3. Remove PBS and add 0.5 ml ice-cold 1X Cell Lysis Buffer plus 1 mM PMSF to each plate (10 cm diameter) and incubate the plate on ice for 5 min.
- 4. Scrape cells off the plate and transfer to an appropriate tube. Keep on ice.
- 5. Sonicate lysates on ice.
- 6. Microcentrifuge for 10 min (x14,000 rpm) at 4°C and transfer the supernatant to a new tube. The supernatant is the cell lysate. Store at -80°C in single-use aliquots.

For suspension cells

- 1. Remove media by low speed centrifugation (\sim 1,200 rpm) when the culture reaches 0.5–1.0 x 10⁶ viable cells/ml.
- Treat cells by adding fresh media containing regulator for desired time.

 2. Collect cells by low speed centrifugation (~1,200 rpm) and wash once with 5–10 ml ice-cold 1X PBS.
- 3. Cells harvested from 50 ml of growth media can be lysed in 2.0 ml of 1X cell lysis buffer plus 1 mM PMSF.
- 4. Sonicate lysates on ice.
- 5. Microcentrifuge for 10 min (x14,000 rpm) at 4°C and transfer the supernatant to a new tube. The supernatant is the cell lysate. Store at -80°C in single-use aliquots.

C. Coating Procedure

- 1. Rinse microplate with 200 μl of dH_2O , discard liquid. Blot on paper towel to make sure wells are dry.
- 2. Dilute capture antibody 1:100 in $1\bar{X}$ PBS. For a single 96 well plate, add 100 μ l of capture antibody stock to 9.9 ml 1X PBS. Mix well and add 100 μ l/well. Cover plate and incubate overnight at 4°C (17–20 hr).
- 3. After overnight coating, gently uncover plate and wash wells:
 - 1. Discard plate contents into a receptacle.
 - 2. Wash four times with wash buffer, 200 µl each time per well. For each wash, strike plates on fresh paper towels hard enough to remove the residual solution in each well, but do not allow wells to completely dry at any time.
 - 3. Clean the underside of all wells with a lint-free tissue.
- 4. Block plates. Add 150 μ l of blocking buffer/well, cover plate, and incubate at 37°C for 2 hr.
- 5. After blocking, wash plate (Section C, Step 3). Plate is ready to use.

D. Test Procedure

1. Lysates can be used undiluted or diluted in blocking buffer. $100~\mu l$ of lysate is added per well. Cover plate and incubate at $37^{\circ}C$ for 2 hr.

- 2. Wash plate (Section C, Step 3).
- 3. Dilute detection antibody 1:100 in blocking buffer. For a single 96 well plate, add 100 µl of detection antibody Stock to 9.9 ml of blocking buffer. Mix well and add 100 µl/well. Cover plate and incubate at 37°C for 1 hr.
- 4. Wash plate (Section C, Step 3).
- 5. Secondary antibody, either streptavidin anti-mouse or anti-rabbit-HRP, is diluted 1:1000 in blocking buffer. For a single 96 well plate, add 10 μ l of secondary antibody stock to 9.99 ml of blocking buffer. Mix well and add 100 μ l/well. Cover and incubate at 37°C for 30 min.
- 6. Wash plate (Section C, Step 3).
 7. Add 100 μl of TMB substrate per well. Cover and incubate at 37°C for 10 min.
- 8. Add 100 µl of STOP solution per well. Shake gently for a few seconds.
- 9. Read plate on a microplate reader at absorbance 450 nm.
 - 1. Visual Determination: Read within 30 min after adding STOP solution.
 - 2. Spectrophotometric Determination: Wipe underside of wells with a lint-free tissue. Read absorbance at 450 nm within 30 min after adding STOP solution.

posted January 2008

revised Sepetember 2013

Orders: 877-616-CELL (2355) • orders@cellsignal.com • Support: 877-678-TECH (8324) • info@cellsignal.com • Web: cellsignal.com For Research Use Only. Not for Use in Diagnostic Procedures.