



Orders: 877-616-CELL (2355)  
orders@cellsignal.com

Support: 877-678-TECH (8324)

Web: info@cellsignal.com  
cellsignal.com

3 Trask Lane | Danvers | Massachusetts | 01923 | USA

# FastScan™ Total Neurogranin ELISA Kit

Store at +4C  
#99828

1 Kit (96 assays)

Species Cross Reactivity: H M R    UniProt ID: #Q92686    Entrez-Gene Id: #4900

**For Research Use Only. Not for Use in Diagnostic Procedures.**

| Product Includes                                        | Product # | Quantity | Color               |
|---------------------------------------------------------|-----------|----------|---------------------|
| FastScan™ ELISA Microwell Strip Plate, 96 Well          | 53257     | 96 tests |                     |
| Neurogranin Rabbit Capture mAb                          | 33414     | 1 ea     | Green (Lyophilized) |
| Neurogranin Rabbit HRP-linked mAb                       | 58057     | 1 ea     | Red (Lyophilized)   |
| FastScan™ ELISA Capture Antibody Diluent                | 16076     | 3 ml     | Green               |
| FastScan™ ELISA HRP Antibody Diluent                    | 28120     | 3 ml     |                     |
| TMB Substrate                                           | 7004      | 11 ml    |                     |
| STOP Solution                                           | 7002      | 11 ml    |                     |
| Sealing Tape                                            | 54503     | 1 ea     |                     |
| ELISA Wash Buffer (20X)                                 | 9801      | 25 ml    |                     |
| FastScan™ ELISA Cell Extraction Buffer (5X)             | 69905     | 10 ml    |                     |
| FastScan™ ELISA Cell Extraction Enhancer Solution (50X) | 25243     | 1 ml     |                     |
| FastScan™ ELISA Kit #99828 Positive Control Type 1      | 96367     | 1 ea     |                     |

Kit contents scale proportionally with size, except sealing tape.

Example: The V1 kit contains 5X the listed quantities above, but will exclude the sealing tape.

The microwell plate is supplied as 12 8-well modules - Each module is designed to break apart for 8 tests.

## Description

The FastScan™ Total Neurogranin ELISA Kit is a sandwich enzyme-linked immunosorbent assay (ELISA) that detects endogenous levels of neurogranin. To perform the assay, sample is incubated with a capture antibody conjugated with a proprietary tag and a second detection antibody linked to HRP, forming a sandwich with neurogranin in solution. This entire complex is immobilized to the plate via an anti-tag antibody. The wells are then washed to remove unbound material. TMB is then added. The magnitude of observed signal is proportional to the quantity of neurogranin.

\*Antibodies in kit are custom formulations specific to kit.

## Specificity/Sensitivity

The FastScan™ Total Neurogranin ELISA Kit detects endogenous levels of neurogranin, as shown in Figure 1. This kit detects proteins from the indicated species, as determined through in-house testing, but may also detect homologous proteins from other species.

## Background

Neurogranin is a small neuronal protein that binds to calmodulin in the absence of Ca<sup>2+</sup> (1,2). Calmodulin binds Ca<sup>2+</sup> and acts as a Ca<sup>2+</sup> sensor to drive Ca<sup>2+</sup>-dependent signal transduction pathways by modulating its interaction with various kinases and phosphatases (3). In the postsynaptic compartment of the neuron, Ca<sup>2+</sup> is carefully regulated and is a critical regulator of synaptic function and plasticity (4). Neurogranin is primarily expressed in the brain and is enriched in somato-dendritic compartments of projection neurons in various regions of the brain (5). Postsynaptically-enriched neurogranin likely influences Ca<sup>2+</sup> or Ca<sup>2+</sup>/calmodulin-dependent neuronal functions, including synaptic plasticity, by binding to and releasing calmodulin in a Ca<sup>2+</sup>-dependent manner (6-9). Genetic variants in the gene encoding neurogranin, *NRGN*, are linked to several neuropsychiatric diseases, including schizophrenia (10). Interestingly, increase in cerebrospinal fluid neurogranin is correlated with neurodegenerative disease progression, which suggests that neurogranin could act as a biomarker for diseases like Alzheimer's disease (11).

## Background References

- Baudier, J. et al. (1991) *J Biol Chem* 266, 229-37.
- Gispén, W.H. et al. (1991) *Mol Neurobiol* 5, 61-85.
- Chin, D. and Means, A.R. (2000) *Trends Cell Biol* 10, 322-8.
- Blackstone, C. and Sheng, M. (2002) *Front Biosci* 7, d872-85.
- Gerendasy, D.D. and Sutcliffe, J.G. (1997) *Mol Neurobiol* 15, 131-63.

6. Huang, K.P. et al. (2004) *J Neurosci* 24, 10660-9.
  7. Pak, J.H. et al. (2000) *Proc Natl Acad Sci USA* 97, 11232-7.
  8. Zhong, L. et al. (2009) *EMBO J* 28, 3027-39.
  9. Jones, K.J. et al. (2018) *Proc Natl Acad Sci USA* 115, E5805-E5814.
  10. Jin, L. et al. (2019) *Medicine (Baltimore)* 98, e18518.
  11. Willemse, E.A.J. et al. (2018) *Clin Chem* 64, 927-37.
- 

## Trademarks and Patents

Cell Signaling Technology is a trademark of Cell Signaling Technology, Inc.

FastScan™ ELISA is a registered trademark of Cell Signaling Technology, Inc.

U.S. Patents 9,086,407, 9,261,500, and 9,476,874, foreign equivalents, and child patents deriving therefrom.

All other trademarks are the property of their respective owners. Visit [cellsignal.com/trademarks](http://cellsignal.com/trademarks) for more information.

## Limited Uses

Except as otherwise expressly agreed in a writing signed by a legally authorized representative of CST, the following terms apply to Products provided by CST, its affiliates or its distributors. Any Customer's terms and conditions that are in addition to, or different from, those contained herein, unless separately accepted in writing by a legally authorized representative of CST, are rejected and are of no force or effect.

Products are labeled with For Research Use Only or a similar labeling statement and have not been approved, cleared, or licensed by the FDA or other regulatory foreign or domestic entity, for any purpose. Customer shall not use any Product for any diagnostic or therapeutic purpose, or otherwise in any manner that conflicts with its labeling statement. Products sold or licensed by CST are provided for Customer as the end-user and solely for research and development uses. Any use of Product for diagnostic, prophylactic or therapeutic purposes, or any purchase of Product for resale (alone or as a component) or other commercial purpose, requires a separate license from CST. Customer shall (a) not sell, license, loan, donate or otherwise transfer or make available any Product to any third party, whether alone or in combination with other materials, or use the Products to manufacture any commercial products, (b) not copy, modify, reverse engineer, decompile, disassemble or otherwise attempt to discover the underlying structure or technology of the Products, or use the Products for the purpose of developing any products or services that would compete with CST products or services, (c) not alter or remove from the Products any trademarks, trade names, logos, patent or copyright notices or markings, (d) use the Products solely in accordance with CST Product Terms of Sale and any applicable documentation, and (e) comply with any license, terms of service or similar agreement with respect to any third party products or services used by Customer in connection with the Products.

#99828

**FastScan™ Total Neurogranin ELISA Kit****FastScan™ ELISA Protocol****A. Solutions and Reagents**

**NOTE:** Prepare solutions with deionized/purified water or equivalent. Prepare only as much reagent as needed on the day of the experiment.

1. **FastScan™ ELISA Microwell Strip Plate, 96 well (#53257):** Bring all to room temperature before opening bag/use. Unused microwell strips should be returned to the original re-sealable bag containing the desiccant pack and stored at 4°C.
2. **1X ELISA Wash Buffer:** Prepare by diluting ELISA Wash Buffer (20X) (included in each kit) to 1X with deionized water.
3. **1X Cell Extraction Buffer:** Prepare by diluting FastScan™ ELISA Cell Extraction Buffer (5X) #69905 and FastScan™ ELISA Cell Extraction Enhancer Solution (50X) #25243\* to 1X with deionized water. This buffer can be stored at 4°C for short-term use (1-2 weeks). To make 10 mL 1X Cell Extraction Buffer, combine 7.8 mL deionized water, 2 mL FastScan™ ELISA Cell Extraction Buffer (5X), and 200 µL FastScan™ ELISA Cell Extraction Enhancer Solution (50X). Alternatively, Enhancer Solution may be added to the Cell Extraction Buffer after extraction of cells or tissue. When using the 1X Cell Extraction Buffer as a sample diluent for the assay, it is recommended to equilibrate it to room temperature prior to use.

**\*IMPORTANT:** The provided FastScan™ ELISA Cell Extraction Enhancer Solution (50X) may precipitate when stored at 4°C. To dissolve, warm briefly at 37°C and mix gently. The FastScan™ ELISA Cell Extraction Enhancer Solution (50X) can be stored at room temperature to avoid precipitation.

**NOTE:** The 1X Cell Extraction Buffer contains phosphatase inhibitors. Protease inhibitors should be added to the 1X Cell Extraction Buffer immediately prior to lysing cells. Additional phosphatase inhibitors can also be added (e.g. Protease/Phosphatase Inhibitor Cocktail (100X) #5872, not supplied).

4. **FastScan™ ELISA Capture Antibody Diluent:** Green diluent for reconstitution of the Capture Antibody.
5. **FastScan™ ELISA HRP Antibody Diluent:** Diluent (amber bottle) for reconstitution of the HRP-linked Antibody. Protect from light.
6. **4X Capture Antibody:** Reconstitute lyophilized Capture Antibody (green colored cake) with 3 mL FastScan™ ELISA Capture Antibody Diluent (green diluent). Incubate at room temperature for 5 min with occasional gentle mixing to fully reconstitute. For best results, use immediately following antibody reconstitution. Unused reconstituted 4X Capture Antibody may be stored for up to 4 weeks at 4°C, although there may be some loss of signal compared to freshly reconstituted antibody.
7. **4X HRP-linked Antibody:** Reconstitute lyophilized HRP-linked Antibody (red colored cake) with 3 mL FastScan™ ELISA HRP Antibody Diluent. Incubate at room temperature for 5 min with occasional gentle mixing to fully reconstitute. For best results, use immediately following antibody reconstitution. Unused reconstituted 4X HRP-linked Antibody may be stored for up to 4 weeks at 4°C protected from light, although there may be some loss of signal compared to freshly reconstituted antibody.
8. **Antibody Cocktail:** Combine equal volumes of the reconstituted 4X Capture and 4X HRP-linked Antibodies immediately prior to assay and mix. To make 6 mL of the Antibody Cocktail (enough for 1x 96-well plate), combine 3 mL 4X Capture Antibody with 3 mL 4X HRP-linked Antibody.
9. **Positive Control:** Reconstitute 1 vial of lyophilized Positive Control (refer to product datasheet or vial label to determine which type of Positive Control is included with the kit):
  - i. For Positive Control Type 1, add 250 µL deionized water.
  - ii. For Positive Control Type 2, add 500 µL 1X Cell Extraction Buffer.

Mix thoroughly and gently, hold at room temperature for 1 minute and then follow the steps outlined below in the "Test Procedure" section. Positive Controls are recommended to be used immediately after reconstituting, however remaining material may be stored at -80°C (there may be some loss of the positive control signal if freeze/thawed). Positive Controls are supplied as a control reagent, not as an absolute quantitation measure.

**NOTE:** A select number of FastScan™ ELISA kits do not contain a positive control, please refer to Product Includes table on the datasheet for a list of included reagents. Should you need support on how to generate a positive control for those kits, contact CST technical support at support@cellsignal.com.

10. **TMB Substrate (#7004):** Bring to room temperature before use.
11. **STOP Solution (#7002):** Bring to room temperature before use.

**B. Preparing Cell Lysates****For adherent cells**

1. Aspirate media when the culture reaches 80-90% confluence.
2. Remove media and rinse cells once with ice-cold 1X PBS.

3. Remove PBS and add 0.5 mL ice-cold 1X Cell Extraction Buffer (recommended to supplement with protease inhibitors and additional phosphatase inhibitors as needed) to each plate (10 cm diameter) and incubate the plate on ice for 5 min.
4. Scrape cells off the plate and transfer to an appropriate tube. Keep on ice.
5. Sonicate lysates on ice.
6. Microcentrifuge for 5 min (x14,000 rpm) at 4°C and transfer the supernatant to a new tube. The supernatant is the cell lysate. Store at –80°C in single-use aliquots.

#### For suspension cells

1. Remove media by low speed centrifugation (~1200 rpm) when the culture reaches 0.5-1.0 x 10<sup>6</sup> viable cells/ml.
2. Wash once with ice-cold 1X PBS.
3. Cells harvested from 50 mL of growth media can be lysed in 2.0 mL of 1X Cell Extraction Buffer (recommended to supplement with protease inhibitors and additional phosphatase inhibitors as needed).
4. Sonicate lysates on ice.
5. Microcentrifuge for 5 min (x14,000 rpm) at 4°C and transfer the supernatant to a new tube. The supernatant is the cell lysate. Store at –80°C in single-use aliquots.

## C. Test Procedure

**NOTE:** Equilibrate all materials and prepared reagents to room temperature prior to running the assay.

1. Prepare all reagents as indicated above (Section A).
2. Samples should be undiluted or diluted with 1X Cell Extraction Buffer to a 2X protein concentration in order to achieve a final 1X protein concentration upon addition of the antibody cocktail. Individual datasheets for each kit provide a sensitivity curve that serves as a reference for selection of an appropriate starting lysate concentration. The sensitivity curve shows typical results across a range of lysate concentration points.
3. Add 50 µL of each sample or Positive Control to the appropriate wells.
4. Add 50 µL of the Antibody Cocktail to each well.
5. Seal the plate with the supplied sealing tape and incubate for 1 hour at room temperature on a plate shaker set to 400 rpm (moderate agitation).
6. Gently remove the tape and wash wells:
  - a. Discard plate contents into a receptacle.
  - b. Wash 3 times\* with 1X ELISA Wash Buffer, 200 µL each time for every well. After each wash, aspirate or decant from wells. Invert the plate and blot it against clean paper towels to remove the residual solution in each well, but do not allow wells to completely dry at any time.
  - c. Clean the underside of all wells with a lint-free tissue.

**\*NOTE:** Certain FastScan™ ELISA Kits may require additional washes at this step. Any requirements for additional washes will be specifically noted in the product “Description” of the kit’s datasheet.

7. Add 100 µL of TMB Substrate to each well. Seal with tape and incubate the plate in the dark for 15 min at room temperature on a plate shaker (400 rpm, moderate agitation) or alternatively for 10 min at 37°C without shaking.
8. Add 100 µL of STOP Solution to each well. Shake gently for a few seconds.

**NOTE:** Initial color of positive reaction is blue, which changes to yellow upon addition of STOP Solution.

9. Read results:
  - a. **Visual Determination:** Read within 30 min after adding STOP Solution.
  - b. **Spectrophotometric Determination:** Wipe underside of wells with a lint-free tissue. Read absorbance at 450 nm within 30 min after adding STOP Solution.

posted May 2018

revised November 2019